


INTERCONNECTING CABLE ASSEMBLY

Daytronic Model 3X78 Carrier Instrument To SensorData Rotary Transformer Torque Sensor



Daytronic
3X78
Carrier
Instrument
Note 4



SensorData
T231
Rotary
Transformer
Torque
Sensor
Note 5

NOTES

- 1) Belden Cable #9728 with 4 pairs of color-coded cable – each pair individually shielded. Shields of each pair isolated by shrink tubing.
- 2) Location of optional junction box with individual isolated connections; Δ.
- 3) Other (power) connections made at this location must be made inside their own junction box.
- 4) Instrument with PC board connector strip that mates with Amphenol 225-21021-103; 3178 – conditioner & amplifier only, 3278 – conditioner, amplifier, & digital indicator, 3378 – conditioner, amplifier, digital indicator, & Hi-Lo limits..
- 5) Sensor with PT02E-10-6P receptacle.